



Ref. Certif. No.

DE 3 - 31807

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

Product	Laboratory Equipment
Name and address of the applicant	<b>MediSonics AG</b> Weststr. 115 8408 Winterthur SWITZERLAND
Name and address of the manufacturer	MediSonics AG Weststr. 115, 8408 Winterthur, SWITZERLAND
Name and address of the factory	MediSonics AG Weststr. 115, 8408 Winterthur, SWITZERLAND
Ratings and principal characteristics	Input 230Vac(P/N)/400Vac(P/P), 50Hz, 9000W, class I, operating water pressure: 2-10 bar.
Trademark / Brand (if any)	MediSonics
Model/type Ref.	MedUSe 60-F SiXi E/M
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016
as shown in the Test Report Ref. No. which forms part of this certificate	713208094

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 113631 0002 Rev. 00

Date, 2022-06-13

( Hendrik Elze )

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service