



EMC-Testcenter AG  
Moosackerstrasse 77  
CH-8105 Regensdorf  
SWITZERLAND

Accredited according to ISO / IEC 17025 by  
Swiss Accreditation Service SAS and IECCE  
STS 0034  
CBTL 0651

Phone +41 44 302 45 00  
E-mail info@emc-testcenter.com  
Http www.emc-testcenter.com



## TEST CERTIFICATE EMCKC3477A TEST-ZERTIFIKAT EMCKC3477A

**Product** : Broad band ultrasonic system for cleaning  
*Produks*

**Trade mark** : MediSonics  
*Handelsmarke*

**Type** : 100-HP5, MedUSe BBUS-5000  
*Typ*

**Manufacturer** : MediSonics AG  
*Hersteller* Weststrasse 115, 8408 Winterthur, Switzerland

**Test results** : Ref. No.: EMCKP3477A  
*Testresultate*

**A sample of this product was tested and found to be in conformity with the following standards:**

*Ein Exemplar dieses Produktes wurde getestet und die Übereinstimmung mit den folgenden Normen festgestellt:*

IEC 60601-1-2: 2014

EN 60601-1-2: 2015

Regensdorf, 17<sup>th</sup> May 2018

EMC-TESTCENTER AG

Max Hunziker  
General Manager Technics / QA